


DRD NanoBlast96™

Pipetting smaller nanoliters with Deska filled Labcon50 disposable tip



Each discrete sample was aspirated in the fine Differential mode and then blown off in the Bulk or Single Piston mode -- contact free -- into a cup of oil, and weighed. The DRD DeSKa (Dead Space Killah) mandril extension eliminates most of the tip air, enabling the Blastoff linearity to extend down to 170 nanoliters (~250 nanoliters is the generally recommended lowest robust volume for this tip in this application).